

CHANGES PROPOSED IN TENDER REF.NO.1671/TAM/NA/2018
BY THE TECHNICAL COMMITTEE AFTER PRE - BID MEETING
HELD ON 16.07.2018

SECTION VI-TECHNICAL SPECIFICATION WITH WARRANTY DETAILS
SCHEDULE - A

300 NRHM ISM&H wings:

Schedule S. No	Name of the Equipment	Existing Specifications with warranty details	Amend to read as
1.	Treadmill (Automatic)	<ul style="list-style-type: none"> • Motor - 1.5 / 3 HP DC • Walking Area - 120 x 40 cm • Incline - 12 Level Auto <p>Warranty: 5 years full comprehensive warranty</p>	<ul style="list-style-type: none"> • Motor: 1.5Hp to 3Hp DC. • Motor control System: Club - rated PWM with fuse less current limit & dynamic short break circuit protection. • Speed: 1-14KM/H. • Incline: 12 Level auto. • Display type: 5 inch LCD windows display. • Computer feedback: Calories, Heart rate, Speed, Time, Distance. • Program: P1 - P9 Preset program. • Deck: 15mm low maintenance prewaxed phenolic coated both sides. • Cushioning: 6 variable - durometer elastomer cushions. • Belt: 3 - ply urethane coated high conductivity. • Running area: 1330 x 430mm. • User weight: 100kgs. • Speed switch buttons on handle bars. • Emergency stop button. • Wheels for

			transportation Warranty: 5 years full comprehensive warranty
--	--	--	--

#

SCHEDULE - C

Purchase of Equipments to 2 co-located wings and 17 Yoga Wellness Centres

Schedule S.No	Name of the Equipment	Existing Specifications with warranty details	Amend to read as
1.	Wax Machine 10 Kg wax	2 kg Capacity, indirect method	2kg capacity, indirect method with 10kg wax

Schedule S.No	Name of the Equipment	Total Quantity	Amend to read as
52	Steam bath cabin (Model IF(SB) - 0192B)	38	17
53	Massage Table	17	38

Schedule - D

Purchase of Equipments to Quality Control Lab Attached to Pharmacy, AAGHIM, Chennai - 106.

Schedule S.No	Name of the Equipment	Existing Specifications with warranty details	Amend to read as		
CHEMISTRY SECTION					
5.	Electric Balance	Readability 0.01mg	Readability 0.1mg 4 digit.		
PHAMACOGNOSY					
2.	Chemical balance	Readability down to 0.01mg, 4dgigits.	Readability down to 0.01mg, 5digits.		
MICROBIOLOGY SECTION					
3.	Plain Incubator	Temperature range not mentioned	Temperature range is 30° c to 100° c.		
5.	Auto valve/Sterilizer	Item Name	Specification with warranty	Item Name	Specification with warranty
		Autovalve/ Sterilizer	Model not mentioned	Autoclave/ Sterilizer	Model: Vertical Type.

Schedule - E

Purchase of Equipments to Drugs Testing Laboratory(I.M) , Chennai - 106.

Schedule S.No	Name of the Equipment	Existing Specifications with warranty details	Amend to read as
1.	Microscope with Digital Camera	<p>Hallogen light illuminator Achromatic objectives, 4 x, 10 x, 40x (spring), 100x (oils spring) Paired eye pises (10x)Antifungal optics Coaxial coarse and fine focusing controls Anti fungus lens. Eye piece. body. Preferred Make: Olympics, Leica or any other equivalent.</p> <p>Warranty: 5 years full comprehensive warranty.</p>	<ul style="list-style-type: none"> • Trinocular Microscope with 6watts LED illumination(with five years warranty for source) • Infinity optical corrected system with Fly Eye optics for Uniform illumination. • Sidentopf-type trinocular tube with light path selector of 100:0, 0:100. • Eyepiece with field of view 22mm and Diopter adjustment facility on both the eyepieces. • Objectives of 4X, 10X, 20X, 40X Flat Field and 100X oil as optional. • Sextuple nosepiece to accommodate 6 objectives at a time. • Right Hand Stage Holder with two specimen holder with refocusing stage mechanism. • Focus: Coarse Stroke - 30mm per rotation, Fine Stroke - 0.1mm per rotation. • Minimum Reading 1 micron, coarse motion torque adjustable, refocusing function and image capture button. • Simple Polarizing

			<p>Attachment with polarizer and Intermediate tube with analyzer.</p> <ul style="list-style-type: none"> • A high - resolution Digital CCD colour camera with 5 Mega Pixels, 2/3" Chip CCD Sensor, pixel size 3.4 x 3.4 microns, 14 bits depth, live display mode: 21 fps(1280 x 960 pixel), 37 fps(1280 x 480 pixel). Exposure time 1/1000 to 600 sec, with C mount adapter 0.7x. • Data processing system i3, 2GB RAM, 500GB HDD, 18.5 Monitor, Keyboard with mouse and UPS. • Warranty: 5 years full comprehensive warranty. <p>Note: Microscope and Digital Colour Camera are to be quoted from single manufactures only.</p>
2.	Electronic Analytical Balance	<p>Self-calibration Capacity: 0.01mg-220gm Pan size: 80 mm Linearity : ± 0.2 mg Repeatability standard deviation: ≥ 0.1 mg Response time : 12/3 sec Sensitivity against temperature change Specific gravity measurement kits.</p> <p>Printer should be provided.</p> <p>Warranty: 5 years full comprehensive warranty</p>	<ul style="list-style-type: none"> • Self - Calibration capacity:0.1mg - 220gm. • Pan size: 80mm. • Linearity: ± 0.2 mg. • Repeatability standard deviation: ≥ 0.1mg. • Response time: 12/3 sec. • Sensitivity against temperature change specific gravity measurement kits. • Display - Backlite LCD. • Data interphase CL and RS to 232 C with

			printer. • Warranty: 5 years full comprehensive warranty.
--	--	--	--

#

Schedule - F

**Purchase of Material and Supplies to Yoga & Naturopathy Centres
(2 co-located wings, 51NRHM centres, 49 Life Style Clinics, Govt Y&N Medical
College, 2 PHC and Govt. Multi Speciality Hospital, Omandurar estate,
Chennai)**

Schedule S.No	Name of the Equipment	Existing		Amend to read as	
		Item Name	Specification with warranty	Item Name	Specification with warranty
28.	Wax	Wax 5	5kg	Wax	1kg x 5

#

Schedule - G

**Purchase of Equipments to Government Siddha Medical College,
Chennai - 106.**

Schedule S.No	Name of the Equipment	Existing	Amend to read as
11.	Hot air oven	For drying slides - 45 degree Celsius	This item is removed.

#

SECTION VIII : QUALIFICATION CRITERIA

S.No:16. The bidders should provide samples and arrange for product demonstration wherever it is required for the product quoted by them.	To be inserted
S. No: 16 to be read as 17.	

#